

National Institute
of Standards and Technology



National Voluntary
Laboratory Accreditation Program

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 1 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

2800 North Lincoln Boulevard
Oklahoma City, OK 73105-4298

Mr. Ken Fraley

Phone: 405-522-5459 Fax: 405-521-4912

E-Mail: kfraley@oda.state.ok.us

URL: <http://www.state.ok.us/lab-bos.htm>

NVLAP Code: 20/A01

ANSI/NCSL Z540-1-1994; Part 1

Compliant

DIMENSIONAL

NVLAP Code: 20/D13

Surveying Rods and Tapes

<i>Range in inches</i>	<i>Best Uncertainty (\pm) in inches^{note 1}</i>	<i>Remarks</i>
1	0.0033	Rigid Rules
2	0.0033	Rigid Rules
3	0.0033	Rigid Rules
4	0.0033	Rigid Rules
5	0.0033	Rigid Rules
6	0.0033	Rigid Rules
7	0.0033	Rigid Rules
8	0.0033	Rigid Rules

December 31, 2004

Effective through

For the National Institute of Standards and Technology

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 2 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

9	0.0033	Rigid Rules
10	0.0033	Rigid Rules
11	0.0033	Rigid Rules
12	0.0033	Rigid Rules
24	0.0057	Rigid Rules
36	0.0081	Rigid Rules
48	0.0104	Rigid Rules

<i>Range in feet</i>	<i>Best Uncertainty (±) in inches^{note 1}</i>	<i>Remarks</i>
1	0.0047	Tapes - Bench Method
2	0.0048	Tapes - Bench Method
3	0.0048	Tapes - Bench Method
4	0.0048	Tapes - Bench Method
5	0.0048	Tapes - Bench Method
6	0.0048	Tapes - Bench Method
7	0.0048	Tapes - Bench Method
8	0.0048	Tapes - Bench Method

December 31, 2004

Effective through

For the National Institute of Standards and Technology

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 3 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

9	0.0049	Tapes - Bench Method
10	0.0049	Tapes - Bench Method
20	0.0091	Tapes - Bench Method
30	0.0133	Tapes - Bench Method
40	0.0174	Tapes - Bench Method
50	0.0216	Tapes - Bench Method
60	0.0258	Tapes - Bench Method
70	0.0299	Tapes - Bench Method
80	0.0341	Tapes - Bench Method
90	0.0383	Tapes - Bench Method
100	0.0424	Tapes - Bench Method
110	0.0466	Tapes - Bench Method
120	0.0507	Tapes - Bench Method
130	0.0549	Tapes - Bench Method
140	0.0591	Tapes - Bench Method
150	0.0632 ^{note 2}	Tapes - Bench Method

December 31, 2004

Effective through



For the National Institute of Standards and Technology

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 4 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

MECHANICAL

NVLAP Code: 20/M08

Mass - Metric

<i>Range</i>	<i>Best Uncertainty (\pm)^{note 1}</i>	<i>Remarks</i>
30 kg	10.3 mg	Echelon I
20 kg	4.6 mg	Echelon I
10 kg	0.7 mg	Echelon I
5 kg	0.42 mg	Echelon I
3 kg	0.35 mg	Echelon I
2 kg	0.33 mg	Echelon I
1 kg	0.066 mg	Echelon I
500 g	0.034 mg	Echelon I
300 g	0.022 mg	Echelon I
200 g	0.015 mg	Echelon I
100 g	0.013 mg	Echelon I
50 g	0.0073 mg	Echelon I
30 g	0.0052 mg	Echelon I

December 31, 2004

Effective through

For the National Institute of Standards and Technology

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 5 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

20 g	0.0044 mg	Echelon I
10 g	0.0046 mg	Echelon I
5 g	0.0024 mg	Echelon I
3 g	0.0016 mg	Echelon I
2 g	0.0012 mg	Echelon I
1 g	0.0011 mg	Echelon I
500 mg	0.00070 mg	Echelon I
300 mg	0.00054 mg	Echelon I
200 mg	0.00050 mg	Echelon I
100 mg	0.00056 mg	Echelon I
50 mg	0.00029 mg	Echelon I
30 mg	0.00019 mg	Echelon I
20 mg	0.00015 mg	Echelon I
10 mg	0.00013 mg	Echelon I
5 mg	0.00010 mg	Echelon I
3 mg	0.00009 mg	Echelon I
2 mg	0.00008 mg	Echelon I

December 31, 2004

Effective through

For the National Institute of Standards and Technology

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 6 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

1 mg	0.00010 mg	Echelon I
1200 kg	10000 mg	Echelon II
750 kg	10000 mg	Echelon II
500 kg	1100 mg	Echelon II
250 kg	480 mg	Echelon II
200 kg	480 mg	Echelon II
100 kg	480 kg	Echelon II
50 kg	64 mg	Echelon II
30 kg	13 mg	Echelon II
20 kg	11 mg	Echelon II
10 kg	7 mg	Echelon II
5 kg	1.8 mg	Echelon II
3 kg	1.3 mg	Echelon II
2 kg	1.1 mg	Echelon II
1 kg	0.26 mg	Echelon II
500 g	0.14 mg	Echelon II
300 g	0.10 mg	Echelon II

December 31, 2004

Effective through



For the National Institute of Standards and Technology

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 7 of 11


CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

200 g	0.08 mg	Echelon II
100 g	0.030 mg	Echelon II
50 g	0.018 mg	Echelon II
30 g	0.014 mg	Echelon II
20 g	0.012 mg	Echelon II
10 g	0.0081 mg	Echelon II
5 g	0.0049 mg	Echelon II
3 g	0.0032 mg	Echelon II
2 g	0.0032 mg	Echelon II
1 g	0.0021 mg	Echelon II
500 mg	0.0018 mg	Echelon II
300 mg	0.0013 mg	Echelon II
200 mg	0.0010 mg	Echelon II
100 mg	0.00089 mg	Echelon II
50 mg	0.00064 mg	Echelon II
30 mg	0.00038 mg	Echelon II
20 mg	0.00044 mg	Echelon II

December 31, 2004



Effective through

For the National Institute of Standards and Technology



ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation

Revised 4/2/04

Page 8 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

10 mg	0.00034 mg	Echelon II
5 mg	0.00030 mg	Echelon II
3 mg	0.00030 mg	Echelon II
2 mg	0.00031 mg	Echelon II
1 mg	0.00026 mg	Echelon II
1200 kg	21000 mg	Echelon III
1000 kg	19000 mg	Echelon III
750 kg	12000 mg	Echelon III
500 kg	3700 mg	Echelon III
300 kg	3300 mg	Echelon III
250 kg	3700 mg	Echelon III
200 kg	2900 mg	Echelon III
100 kg	1800 mg	Echelon III
50 kg	1400 mg	Echelon III
30 kg	180 mg	Echelon III
25 kg	170 mg	Echelon III
20 kg	160 mg	Echelon III

December 31, 2004

Effective through

For the National Institute of Standards and Technology

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 9 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

10 kg	150 mg	Echelon III
5 kg	13.0 mg	Echelon III
3 kg	7.8 mg	Echelon III
2 kg	7.2 mg	Echelon III
1 kg	6.1 mg	Echelon III
500 g	6.0 mg	Echelon III
300 g	3.8 mg	Echelon III
200 g	1.1 mg	Echelon III
100 g	0.50 mg	Echelon III
50 g	0.30 mg	Echelon III
30 g	0.28 mg	Echelon III
20 g	0.28 mg	Echelon III
10 g	0.14 mg	Echelon III
5 g	0.14 mg	Echelon III
3 g	0.14 mg	Echelon III
2 g	0.14 mg	Echelon III
1 g	0.12 mg	Echelon III

December 31, 2004

Effective through

For the National Institute of Standards and Technology

National Institute
of Standards and Technology



National Voluntary
Laboratory Accreditation Program

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 10 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

500 mg	0.12 mg	Echelon III
300 mg	0.12 mg	Echelon III
200 mg	0.077 mg	Echelon III
100 mg	0.070 mg	Echelon III
50 mg	0.052 mg	Echelon III
30 mg	0.052 mg	Echelon III
20 mg	0.052 mg	Echelon III
10 mg	0.055 mg	Echelon III
5 mg	0.048 mg	Echelon III
3 mg	0.048 mg	Echelon III
2 mg	0.048 mg	Echelon III
1 mg	0.040 mg	Echelon III

December 31, 2004

A handwritten signature in black ink, appearing to read 'William R. Mohr'.

Effective through

For the National Institute of Standards and Technology

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 11 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

NVLAP Code: 20/M12

Volume

<i>Range</i>	<i>Best Uncertainty (\pm)^{note 1}</i>	<i>Remarks</i>
5 gal	0.14 in ³	Volume Transfer Method
10 gal	0.24 in ³	Volume Transfer Method
15 gal	0.30 in ³	Volume Transfer Method
20 gal	0.42 in ³	Volume Transfer Method
25 gal	1.0 in ³	Volume Transfer Method
30 gal	1.6 in ³	Volume Transfer Method
50 gal	2.2 in ³	Volume Transfer Method
100 gal	3.5 in ³	Volume Transfer Method
200 gal	6.7 in ³	Volume Transfer Method
375 gal	12 in ³	Volume Transfer Method
5 gal	0.05 in ³	Volume Gravimetric Method
25 gal	0.75 in ³	Volume Gravimetric Method

1. Represents an expanded uncertainty using a coverage factor, $k=2$.

December 31, 2004

Effective through

For the National Institute of Standards and Technology

National Institute
of Standards and Technology



National Voluntary
Laboratory Accreditation Program

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Revised 4/2/04

Page 12 of 11

CALIBRATION LABORATORIES

NVLAP LAB CODE 200396-0

OKLAHOMA BUREAU OF STANDARDS

2. Tapes Greater than 150 feet uncertainty equals $0.0632 + 0.046$ inches per 10 foot interval

December 31, 2004

A handwritten signature in black ink, appearing to read 'William R. Muhl'.

Effective through

For the National Institute of Standards and Technology